

Studies of a Bulk Micromegas using the Cornell/Purdue TPC

Cornell University

T. Anous

R. S. Galik

D. P. Peterson

Purdue University

K. Arndt

G. Bolla

I. P. J. Shipsey

The “Bulk Micromegas”, was prepared
on one of our pad boards by Paul Colas’ group.

This project is supported by
the US National Science Foundation (LEPP cooperative agreement)
and by the US Department of Energy (Purdue HEP group base grant)
and an LCDRD consortium grant (NSF and DoE)

Topics

- Description of the chamber (mostly repeat, a few updates)
- Measurements of the Bulk Micromegas , B=0, Ar-isoC₄H₁₀(7%)
running conditions (training, sparking)
anode signal width
spatial resolution
- Comments on continued preparations for ion feedback measurements

Further information available at the web site:

http://www.lepp.cornell.edu/~dpp/tpc_test_lab_info.html


* presentation at ECFA Valencia	07-November-2006	electron and ion transmission
* presentation at ALCPG Vancouver	18-July-2006	demonstration of ion signal
* presentation at Berkeley TPC Workshop	08-April-2006	Purdue-3M Micromegas
* presentation at ECFA 2005 Vienna	24-November-2005	
* presentation at ALCPG Snowmass	23-August-2005	
* presentation at LCWS05, Stanford	21-March-2005	

TPC

14.6 cm ID field cage - accommodates a 10 cm gas amplification device
64 cm drift field length
22.2 cm OD outer structure (8.75 inch)

“field cage termination” and “final” return lines for the field cage HV distribution allow adjustment of the termination bias voltage with an external resistor.

Read-out end:
field cage termination
**readout pad and
gas amplification module**
pad biasing boards
CLEO II cathode preamps



Electronics

High voltage system:


- 20 kV module
- 2 kV module, 4 channels
- +2 kV module, 4 channels
- +4 kV module, for 3-GEM

Readout:


- VME crate
- PC interface card
- LabView

Struck FADC

- 56 channels (increasing to 88)
- 105 M Hz
- 14 bit
- +/- 200 mV input range
(least count is 0.025mV)
- NIM external trigger input
- circular memory buffer



TPC pad board



Pad board with 2 mm pads.

80 pads on the board

4 layers of 2mm pads


5 layer of 5mm pads
for track definition

For this data set,
limited to 56 channels


use 6 layers:
3 @ 2 mm width
3 @ 5 mm width.

*Resolution measurements
are derived from the
difference in residuals on
adjacent 2mm pad rows.*

Micromegas amplification




10 cm



The “bulk Micromegas”, was prepared on one of our pad boards by Paul Colas’ group. Measurements with the Purdue-3M Micromegas were shown at Vancouver 2006.

Micromegas amplification




The Micromegas is located 0.78 cm from the field cage termination.

HV is distributed to the pads; note blocking capacitors, HV resistors.


Low voltage signals routed to preamps outside (on ribbon cable).

Micromegas is at ground; pads at +410V for Ar-isoC₄H₁₀ (7%).

bulk Micromegas event




Ar-isoC₄H₁₀ (7%) , 200V/cm
Micromegas: 410V / 50 μm



25 MHz , 40 ns
2048 time buckets (81.92 μs)

bulk Micromegas event





Notice that in these events, there is an opposite signal on every channel.

This will affect the apparent charge width.

This "opposite signal" is not unique to this type of Micromegas.



Purdue-3M Micromegas



← 10 cm →

Measurements with the Purdue-3M Micromegas were shown at Vancouver 2006.

A similar “opposite signal” was observed with this device (below).

Purdue-3M Micromegas


Micromegas is commercially made by the 3M corporation in a proprietary subtractive process starting with copper clad Kapton.

Holes are etched in the copper
70 μm spacing
35 μm diameter

Copper thickness: 9 μm

Pillars: remains of etched Kapton.
50 μm height
300 μm diameter at base
1 mm spacing, square array


The shiny surface of the pillars is due to charge build-up from the electron microscope.




Title: Copper Electrodes
Comment: Kirk Arndt

Date: 03-22-2004 Time: 14:57
Filename: PHYSICS2.TIF

Purdue-3M Micromegas event




Ar-isoC₄H₁₀ (10%) , 200V/cm
Micromegas: 410V / 50 mm



25 MHz , 40 ns
2048 time buckets (81.92 μs)

"opposite signal" not in FCT



The "opposite signal" is observed in both the Bulk Micromegas and the Purdue-3M Micromegas.


It is not clear if this signal is coming from the pad board or from the pickup in the electronics.

Shown at left is an event with 8 FADC channels connected to the field cage termination. The preamps are connected to power supplies in the same way.

By not seeing the "opposite signal" in the FCT, electronic pickup is ruled out; The "opposite signal" originates at the pads.

25 MHz , 40 ns
2048 time buckets (81.92 μ s)

Micromegas sparking



Training:

air 500-520V 24 hours

Ar CO₂ 410V 1 hour
 420V 18 hours
 430V 4 days

Ar-isoC₄H₁₀ (7%)

400V / 50 μm 22 hours
 410V 6 days

Non-destructive sparking observed:
 PH ~ 100x typical min. ionizing.
Sparking is picked-up by the scintillator/trigger (pad signal in channel 90±1).

Rate: 7.6 / hour at beginning of Ar CO₂ running

5.9 / hour at beginning and end of Ar-isoC₄H₁₀.

25 MHz , 40 ns
 2048 time buckets (81.92 μs)

Drift velocity / Gain

Drift velocities for various gas mixtures are shown at right (from various sources).

For Ar-isoC₄H₁₀ (7%), expect ~39 mm/μs .



Observed time for a maximum drift 64.7 cm is (410 FADC time buckets)x(40ns/bucket), or 39.5 mm/μs .

The gain for various gas mixtures are shown at right. Sources are indicated.



Although it is difficult to extrapolate for Ar-isoC₄H₁₀ (7%) , at 410V, the gain is about estimated to be ~10⁵ .

While Gain estimates were stated for the Purdue-3M Micromegas at Berkeley, April 2006, the absolute gain requires more study.

However, the gain ratio, Bulk/Purdue, is ~20% .



Charge width / diffusion



The charge width is determined from the fraction of the total charge in 1, 2 or 3 pads, shown above, assuming a gaussian charge distribution.

(The measurement deviates for the 1 and 2 pad measurement at large drift distance. Possibly, the fraction of the signal in a "small" width is overestimated by selecting the maximum.)

The line at left indicates a diffusion constant of $D = .0415 \text{ cm} / (\text{cm})^{1/2}$.
(Recall that this will be affected by the loss of small signals due to the "opposite signal".)

hit resolution (2mm pad)

find tracks

require time coincident signals in 5 layers

there are 6 layers available:

3x 5mm-pad layers,

a single 2mm-pad layer, a 2mm-pad pair

find PH center using maximum PH pad
plus nearest neighbors
(total 2 to 4 pads)

fit, deweighting the 5mm pad measurements

point measurement

low drift (narrow pad distribution function)


hits are corrected for an “effective pad center”

(This is not ideal, but it is what we are currently using.)


resolution difference

RMS of difference in residual
for the adjacent 2mm layers

correct with : $\sigma = \text{RMS} / \sqrt{2}$



Here, the containment width of the pad distribution function is small; any sharing indicates that the charge center of each pad is not the geometric center. Thus, there is a shift of the effective pad center.



cuts, calibration

slope < 0.05

the trigger allows ~ 0.08

|x| < 11 mm

removes poorly measured edge tracks

residual in the single (2mm) layer < 0.4 mm

requires consistent hits in adjacent 5mm layers
although it is higher weighted in the fit


fraction of signal in 1 pad < 99%

much looser than previous analysis

(for low drift bins)

fraction of signal in 2 bins > 80%

removes a type of noise event with
equal pulse height in all pads.



Pad-to-pad pulse height calibration (as large as $\pm \sim 30\%$)

Hit resolution

Fit to $\sigma = (\sigma_0^2 + D^2/n x)^{1/2}$

use $D = .0415 \text{ cm}/(\text{cm})^{1/2}$.


result: $n = 17.4 \pm .5$
 $\sigma_0 = 53 \pm 36 \text{ } \mu\text{m}$
 $\chi^2/\text{dof} = 1.7$

All points are in the fit.


A systematic uncertainty in σ_0 arises from a possible error in determining the time for drift=0.

If T_0 is actually in the center of the first drift bin, then

$\sigma_0 \text{ (modified } T_0) = 103 \text{ } \mu\text{m}$.




Ion Feedback Detection




We continue plans to measure positive ion feed-back into the field cage

using a technique of ion collection, for individual tracks, on the (double) field cage termination.



The method differs from that used by Saclay/Orsay on MicroMegas and by Aachen on GEM. For those measurements, a source was used to create ionization. Current was measured on the cathode.



The ion collection was demonstrated in earlier talks, using a constant bias on the field cage termination plane.




Ion Feedback measurement, with pulsed field cage termination



More sensitive measurements will require a pulsed bias on the field cage termination to provide full electron transmission and full ion collection.

The pulsed bias will require new gated preamplifiers.

These have been assembled and are awaiting testing.



Summary, outlook

We have made measurements of the Bulk Micromegas.

Plan to repeat measurements of the Purdue-3M Micromegas with consistent conditions.

We plan to study a triple-GEM.

We are continuing plans for comparative measurements of ion feed-back. (graduate student)

CLEO will end data taking April 2008 (after 28.5 years).
Cornell proposals to reconfigure CESR for studies of a wiggler-dominated damping ring.

If this proposal is funded,
the CLEO drift chamber will be removed from solenoid as part of the CESR reconfiguration.

In that case, we will be able to run the small prototype in the 1.5 Tesla CLEO magnet, for resolution, and GEM ion/electron transmission studies. (4 weeks /year, maximum)

